

## 11. Assay Procedure Continued...

- Using a repeating dispenser, rapidly dispense 100µl of TMB Substrate (**Reagent 4**) into each well. Incubate the plate for **10 minutes**.
- Add 100µl of Stop Solution (**Reagent 5**) to each well. To allow equal reaction times, the Stop Solution should be added to the wells in the same order as the TMB Substrate.
- Read the optical density (OD) of each well at 450nm in a microplate reader within 10 minutes. A 620nm filter may be used as a reference wavelength.

## 12. Quality Control

Quality control data is supplied on the lot-specific QC certificate included in the kit.

The Positive Control is intended to monitor for substantial reagent failure.

Any well positive by spectrophotometer but without visible colour should be cleaned on the underside and re-read. If OD values below zero are observed, the wavelengths used should be verified, the reader re-blanked to air and the measurements repeated.

## 13. Interpretation of Results

### Quantitative Results

Plot the OD of each standard against its concentration and draw the best-fit curve through the points. Read the unknowns off this curve. Values below 2.5 U/ml are normal. Values above 4U/ml are significantly elevated. Values between 2.5 U/ml and 4 U/ml are borderline. These data are based on studies at Genesis Diagnostics. We recommend that each laboratory establish its own normal range.

Values above 100 should be repeated at a higher dilution e.g. 1:200

### Qualitative Results

Samples with an OD greater than the 2.5 U/ml standard are considered positive for antibodies to Goodpasture's antigen.

## 14. Limitations of the Procedure

Results of this assay should be used in conjunction with clinical findings.

## 15. Performance Characteristics

### Clinical Study

A panel of sera from patients with different pathologies and from healthy blood transfusion donors was studied for anti-GBM antibodies.

Disease state	Anti-GBM >2.5U/ml (+)	Anti-GBM <2.5U/ml (-)
Goodpasture's syndrome	31	0
Rheumatoid Arthritis	0	10
Active SLE	0	20
Vasculitis	0	25
Normal Donors	0	57

Sensitivity: 100%

Specificity: 100%

## 16. Reproducibility

### Within Assay Precision

	n	OD	SD	CV%
Normal Sample	10	0.251	0.018	7.1
Elevated Sample	10	1.112	0.039	3.5

### Between Assay Precision

CV%: < 12%

## 17. Method Summary

- Dilute sera 1:50 with Sample Diluent (**Reagent 1**)
- For quantitative assays, dispense Standards, the Positive Control and the diluted sample into the microplate wells. For qualitative assays, dispense the 2.5 U/ml Standard only.
- Incubate for **30 minutes** at room temperature.
- Wash the wells three times*
- Dispense 100µl of Conjugate (**Reagent 3**) into each well
- Incubate at room temperature for **30 minutes**
- Wash the wells four times*
- Add 100µl of TMB Substrate (**Reagent 4**) to each well
- Incubate at room temperature for **10 minutes**
- Add 100µl Stop Solution (**Reagent 5**) to each well
- Read the optical density at 450nm (single wavelength) or 450/620nm (dual wavelength).

## 18. Further Reading

Bosch X et al, (1991) Clin Nephrol 36, 107-113  
Bowman C. Et al (1983) J Immunol Methods 61, 325 - 333  
Bowman C., et al (1985) J Clin Lab Immunol 17, 197 - 202  
Hellmark, T. et al (1994) Kidney Int Sep 94 823-829  
Jayne D R W et al, (1990) Kidney Int. 37, 965-970  
Lockwood C M et al, (1976) Lancet i:711-715  
Lerner R A et al (1967) J Exp Med. 126, 98  
Neilson, EG. Et al (1993) J Biol Chem 268, 8402-8405. Short A K et al, (1993), Adv Exp Med Biol. 33, 441-444

Eden Research Park, Henry Crabb Road, Littleport, Cambridgeshire,  
CB6 1SE, UK Tel+ 44(0)1353 862220 Fax+44(0)1353 863330  
Email: support@elisa.co.uk Web: www.omegadiagnostics.com  
Certified to ISO9001:2008, ISO13485:2003

Genesis Diagnostics Ltd is a subsidiary of Omega Diagnostics Group PLC

G • E • N • E • S • I • S  
*Diagnostics*

## GBM IgG ELISA Kit

Quantitative/qualitative assay for glomerular basement membrane IgG antibodies

Product Code: GD004 / GD004A

For *in vitro* Diagnostic Use



### 1. Materials Included in the Kit

- Microplate:** 12 X 8 break-apart strips (96 tests) or 6x8 break apart strips (48 tests) pre-coated with Goodpasture's antigen, with holder in a foil bag with desiccant.
- Reagent 1: Sample Diluent** 150mM Tris-buffered saline, pH 7.2 with antimicrobial agent, 10ml, (blue), **concentrate** (x15)
- Reagent 2: Wash Buffer** 100mM Tris-buffered saline with detergent, pH 7.2, 100 ml, **concentrate** (x10)
- Reagent 3: Conjugate** rabbit anti-human IgG conjugated to horseradish peroxidase in protein stabilising solution and antimicrobial agent, 12 ml, (red), ready to use
- Reagent 4: TMB Substrate** aqueous solution of TMB and hydrogen peroxide, 12 ml, ready to use
- Reagent 5: Stop Solution** 0.25M sulphuric acid, 12 ml, ready to use
- Standards:** 2.5, 5, 10, 25, 50 & 100 U/ml, 2 ml (96 test) or 1 ml (48-test) of 10mM Tris-buffered saline containing human serum IgG/ antibodies to the Goodpasture's antigen, ready to use
- Positive control:** 2 ml (96 test) or 1 ml (48-test) of 10mM Tris-buffered saline containing human serum antibodies to Goodpasture's antigen, ready to use
- Instructions for use**

## 2. Other Equipment Required

Test tubes for dilution • graduated cylinder for preparing wash buffer • precision pipettes and disposable tips to deliver 10µl, 100µl, 1ml • EIA microplate washer or multi-channel pipette or wash bottle • distilled or de-ionised water • absorbent paper • EIA microplate reader with 450nm and optional 620nm reference filter. Alternatively, a suitable, self-validated automated system may be used.

Instrumentation, whether manual or automated, should meet the following criteria: pipettes with better than 3% imprecision with no carry over between pipetting steps; microplate washers should remove 99% of fluid; automated machines should minimise time between washing and adding the next reagent.

## 3. Intended Use

The GBM IgG kit is a rapid ELISA method for the detection of IgG antibodies to Goodpasture's antigen present in the glomerular basement membrane (GBM). The kit is intended for use in clinical laboratories as an aid to the diagnosis of Goodpasture's syndrome. The components of the kit are for in vitro diagnostic use only.

## 4. Explanation of the Test

Antibodies to the GBM are associated with autoimmune glomerulonephritis with or without lung haemorrhage.

Goodpasture's syndrome is a reno-pulmonary syndrome characterised by a rapidly progressive glomerulonephritis, which, if undiagnosed, may develop into a severe and often-fatal lung haemorrhage. Autoantibodies associated with the disease are directed against the Goodpasture's antigen, which has been localised to the NC1 domain of the alpha-3 chain of type-IV collagen in the GBM. Less than one third of patients with reno-pulmonary syndromes have antibodies against the GBM: the majority have either PR-3 ANCA or MPO-ANCA.

Early recognition of Goodpasture's syndrome is essential for successful treatment, which involves plasma exchange to remove anti-GBM antibody. The Genesis GBM IgG ELISA kit uses microplates coated with purified antigen derived from the NC1 domain of the alpha-3 chain of type-IV collagen and is specific for the detection of autoantibodies to the Goodpasture antigen. The GBM IgG ELISA is a sensitive assay of value in early disease detection, in monitoring therapy and in the detection of disease re-occurrence following renal transplantation. The assay may be used quantitatively or qualitatively.

## 5. Principle of the Test

Diluted serum samples are incubated with purified Goodpasture's antigen immobilised on microtitre wells. After washing away unbound serum components, rabbit anti-human IgG conjugated to horseradish peroxidase is added to the wells, and this binds to surface-bound antibodies in the second incubation. Unbound conjugate is removed by washing, and a solution containing 3,3',5,5'-tetramethylbenzidine (TMB) and enzyme substrate is added to trace specific antibody binding. Addition of Stop Solution terminates the reaction and provides the appropriate pH for colour development. The optical densities of the standard(s), positive control and samples are measured using a microplate reader at 450nm.

## 6. Safety Precautions

1. All reagents in this kit are for *in vitro* diagnostic use only.
2. Only experienced laboratory personnel should use this test. The test protocol must be followed strictly.
3. CAUTION: the device contains material of human and animal origin and should be handled as a potential transmitter of diseases. All human source material used in the preparation of standards and control for this product have been tested and found negative by ELISA for antibodies to HIV, HbsAg and HCV. No test method, however, can offer complete assurance that infectious agents are absent. Therefore, all reagents containing human material should be handled as if potentially infectious. Operators should wear gloves and protective clothing when handling any patient sera or serum based products.
4. Reagents of this kit contain antimicrobial agents and the Substrate solution contains 3,3',5,5'-tetramethylbenzidine. Avoid contact with the skin and eyes. Rinse immediately with plenty of water if any contact occurs.
5. The Stop Solution contains 0.25M sulphuric acid. Avoid contact with skin and eyes. Rinse immediately with plenty of water if contact occurs.
6. Any liquid that has been brought into contact with potentially infectious material has to be discarded in a container with a disinfectant. Dispose of plates and specimens as clinical waste. Any unused reagents should be flushed away with copious amounts of water. Disposal must be performed in accordance with local legislation.

## 7. Technical Precautions

1. Strips and solutions should not be used if the foil bag is damaged or liquids have leaked.
2. Allow all reagents and the microplate to reach room temperature before use. Ensure that the microplate foil bag containing any unused strips is well sealed and contains the desiccant to avoid moisture. Store at 2 – 8°C after use.
3. The sample diluent X15 concentrate contains 0.09% sodium azide as preservative. Prepare sufficient working strength diluent for the assay run. However, if the working strength diluent is to be stored for more than 1 week, add sodium azide (0.9g/L). Store unused sample diluent concentrate and dilute sample diluent at 2 – 8°C.
4. When automating, consider excess volumes required for setting up the instrument and dead volume of robot pipette
5. Include the Positive Control in every test run to monitor for reagent stability and correct assay performance.
6. Strictly observe the indicated incubation times and temperature.
7. Ensure that no cross-contamination occurs between wells. Keep all pipettes and other equipment used for Conjugate completely separate from TMB Substrate.
8. When pipetting Conjugate or TMB Substrate, aliquots for the required numbers of wells should be taken to avoid multiple entry of pipette tips into the reagent bottles. Never pour unused reagents back into the original bottles.
9. Do not allow microwells to dry between incubation steps.
10. Strictly follow the described wash procedure. Insufficient washing may cause high background signal.
11. Avoid direct sunlight and exposure to heat sources during all incubation steps.
12. Replace colour-coded caps on their correct vials to avoid cross-contamination
13. It is important to dispense all samples and the positive control into the wells without delay. Therefore ensure that all samples are ready to dispense.

## 8. Shelf Life and Storage Conditions

On arrival, store the kit at 2 - 8°C. Once opened the kit is stable for 3 months (or until its expiry date if less than 3 months). Do not use kits beyond their expiry date. Do not freeze any kit component. The diluted Wash Buffer and Sample Diluent (see Technical Precautions) have a shelf life of 3 months if stored in a closed bottle at 2 - 8°C.

## 9. Specimen Collection and Storage

Serum or plasma samples may be used and should be stored at -20°C for long-term storage. Frozen samples must be mixed well after thawing and prior to testing. Repeated freezing and thawing can affect results. Addition of preservatives to the serum sample may adversely affect the results. Microbially contaminated, heat-treated or specimens containing particulate matter should not be used. Grossly haemolysed, icteric or lipaemic specimens should be avoided.

## 10. Preparation of Reagents

1. Dilute the Sample Diluent (**Reagent 1**) 1:14 in distilled water to make sufficient buffer for the assay run e.g. add 10ml sample diluent concentrate to 140 ml water.
2. Dilute the Wash Buffer (**Reagent 2**) 1:9 in distilled water to make sufficient buffer for the assay run e.g. add 50ml wash buffer concentrate to 450ml water.

## 11. Assay Procedure

1. Dilute patient samples 1:50 in diluted Sample Diluent (e.g. 20µl serum plus 1ml diluent).
2. Assemble the number of strips required for the assay.
3. For quantitative assays, dispense 100 µl of sample diluent as the 0 U/ml standard, followed by 100 µl of each of the remaining Standards, the Positive Control and the diluted patient samples into appropriate wells. For qualitative assays, dispense only the 2.5 U/ml Standard together with the positive control and samples.
4. Incubate for **30 minutes** at room temperature.
5. After 30 minutes, decant or aspirate the well contents and wash the wells 3 times using automated washing or the manual wash procedure (see below). Careful washing is the key to good results. **Do not allow the wells to dry out.**  
Manual Wash Procedure  
Empty the wells by inversion. Using a multi-channel pipette or wash bottle, fill the wells with Wash Buffer. Empty by inversion and blot the wells on absorbent paper. Repeat this wash process 2 more times.
6. Dispense 100µl of Conjugate (**Reagent 3**) into each well. Incubate the wells for **30 minutes** at room temperature.
7. After 30 minutes, discard the well contents and carefully wash the wells 4 times with Wash Buffer. Ensure that the wells are empty but do not allow to dry out.